



DATA SHEET



Size	Reorder#	Dimension		
		Target weight (g/pcs)	Palm Width (mm)	Length (mm)
5.5	C3655-96	9.6 ± 0.5	72 ± 4	min. 278
6.0	C3660-96	10.2 ± 0.5	77 ± 5	min. 280
6.5	C3665-96	10.8 ± 0.5	83 ± 5	min. 280
7.0	C3670-96	11.4 ± 0.5	89 ± 5	min. 283
7.5	C3675-96	12.0 ± 0.5	95 ± 5	min. 287
8.0	C3680-96	12.6 ± 0.5	102 ± 6	min. 288
8.5	C3685-96	13.2 ± 0.5	108 ± 6	min. 290
9.0	C3690-96	13.8 ± 0.5	114 ± 6	min. 290
<i>*non-release criteria</i>				
Product Part #		362xx.110122035		
510K & MDL number		K020019 & D036512		
Design and Feature		Hand specific, curved fingers, micro roughened surface and beaded cuff		
Type		Powder free, polymer coated and Gamma-sterilized surgical gloves		
Material		Natural Rubber Latex		
Colour		Natural		
Surface Treatment		Polymer coated on inner surface to ease donning		
Powder Residue (mg/glove)		Max. 2		
Protein Content (µg/dm ²)		≤ 50		
Single-wall Thickness (mm)	Finger	Specification		
	Palm	0.21 ± 0.03		
	Cuff	0.17 ± 0.03		
<i>*All sizes</i>		0.15 ± 0.02		
Physical Properties	Tensile Strength (MPa)	Unaged	Aged	
	Ultimate Elongation (%)	Min. 24	Min. 18	
	Stress at 500% (MPa)	Min. 750	Min. 560	
	Force at break (N)	Max. 5.5	N/A	
		≥ 9	≥ 9	
Packing Mode		1 pairs of gloves per inner wrap 1 inner wrap per pouch 50 pouch per dispenser 4 dispenser per carton		
Glove Marking		The cuff of each glove is marked with 'WRP', 'Hand and Size'		
Lot # Identification on Finished Goods		Lot Number Structure : YMMPPPPSS (9 Digits) Y = Year of packing (e.g. 2 for year 2022, 3 for year 2023, etc.) MM = Month of packing (e.g. 09 for Sept., 10 for Oct., etc.) PPPP = WRP's Packing Work Order# (PWO) SS = Size (00=XS, 01=S, 02=M, 03=L, 04=XL and 05=XXL)		
Shelf Life Claim		3 years upon manufacturing date		
Pre-shipment Inspection	Dimension	N=13, Median		
	Physical Properties	N=13, Median		
	1000ml Water Leak	G-1, AQL 0.65		
	Protein Content	N=8		
	Powder Free Residue	N=3		
	Major Visual Inspection	G-1, AQL 2.5		
	Minor Visual Inspection	G-1, AQL 4.0		
Product Conformance		<ul style="list-style-type: none"> Medical Device: in compliance with 93/42/EEC EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018 Type B, EN ISO 374-2:2019, EN ISO 374-4:2019, EN ISO 374-5:2016, EN 16523-1:2015+A1:2018, CE2797 EN 455 Part 1/2/3/4 ASTM D3577 		
Quality Assurance		<ul style="list-style-type: none"> MDSAP ISO 13485:2016 Quality Management System ISO 9001:2015 Quality Management System EN ISO 13485:2016 		

Note: The above information is a guideline of typical performance values and characteristic of the product.

WRP ASIA PACIFIC SDN. BHD.



Name : Rosnaini Binti Rosjidda
Title : Product Management Manager