



DATA SHEET



NITRILE EXAMINATION GLOVES ULTRA PLUS

NITRILE EXAMINATION GLOVES, POWDER FREE, NON-STERILE

Reorder #	Size: : XS S M L XL	D1400-28 D1401-28 D1402-28 D1403-28 D1404-28	
Product Part #	142xx.102511030		
510K # / MDL #	N/A		
Design and Feature	Ambidextrous, Finger textured (E4) and beaded cuff		
Type	Powder free and non-sterile examination gloves		
Material	100% nitrile (Acrylonitrile-butadiene)		
Color	Blue (PMS285)		
Surface Treatment	Chlorination on donning side		
Powder Free Residue (mg/glove)	≤ 2		
Dimension (mm)	Size: : XS S M L XL	<u>Palm Width</u> ≤ 80 80 ± 10 95 ± 10 110 ± 10 ≥ 110	<u>Length</u> Min. 270 Min. 270 Min. 270 Min. 270 Min. 270
Single-wall Thickness (mm) *All sizes	Finger Palm Cuff	<u>Spec</u> 0.11 ± 0.02 0.08 ± 0.02 0.07 ± 0.02	
Physical Properties	Tensile Strength (MPa) Ultimate Elongation (%) Stress at 500% (MPa) Force at Break (N)	<u>Before Aging</u> Min. 14 Min. 500 N/A ≥ 6	<u>After Aging</u> Min. 14 Min. 400 N/A ≥ 6
Packing Mode	200 gloves per dispenser (XS-L) 180 gloves per dispenser (XL) 10 dispensers per carton /shipping case		
Glove Marking	No marking.		
Lot # identification on Finished Goods	Lot Number Structure: YMMPPPPSS (9 digits) Y = Year of Packing MM = Month of Packing PPPP = WRP's Packing Work Order # (PWO) SS = Size (00=XS, 01=S, 02=M, 03=L and 04=XL)		
Recommended Shelf Life	5 years upon manufactured date.		
Pre-shipment Inspection *Single-Normal Sampling Plan	Dimension Physical Properties 1000ml Water Leak Powder Free Residue Major Visual Inspection Minor Visual Inspection	<u>EN Standard</u> N=13, Median; S-2, AQL 4.0 N=13, Median; S-2, AQL 4.0 G-I, AQL 1.5 N=5 G-I, AQL 2.5 G-I, AQL 4.0	
Product Conformance	<ul style="list-style-type: none"> Personal Protective Equipment of Complex Design Category III, in compliance with European Regulation (EU) 2016/425, tested to EN 420:2003+A1:2009, EN ISO 374-1:2016 Type B, EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013 & EN ISO 374-5:2016, CE2797; CE Class I Medical Device, in compliance with MDR (EU) 2017/745, type tested to EN 455-1:2000, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009 ASTM D6319 		
Quality Assurance	<ul style="list-style-type: none"> US FDA Quality System Regulation (QSR) ISO9001:2015 Quality Management Systems EN ISO13485:2016 and ISO13485:2016 Quality Management Systems 		

Note: The above information is a guideline of typical performance values and characteristic of the product; and not to be used as actual product specification.